

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address COMMISSIONER FOR PATENTS FO Box 1430 Alexandria, Virginia 22313-1450 www.tepto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/785,082	02/25/2004	Jessica E. Barzilai	4485C1US	8544
50000 - 7			EXAMINER	
			SIEFKE, SAMUEL P	
			ART UNIT	PAPER NUMBER
	,		1797	
			MAIL DATE	DELIVERY MODE
			06/03/2010	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

## Application No. Applicant(s) 10/785.082 BARZILAI ET AL. Office Action Summary Examiner Art Unit SAM P. SIEFKE 1797 -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --Period for Reply A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS. WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION. - Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication. If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication - Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b). Status Responsive to communication(s) filed on 2a) ☐ This action is FINAL. 2b) This action is non-final. 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213. Disposition of Claims Claim(s) is/are pending in the application. 4a) Of the above claim(s) is/are withdrawn from consideration. 5) Claim(s) \_\_\_\_\_ is/are allowed. 6) Claim(s) 49-51, 53-56 and 58-70 is/are rejected. 7) Claim(s) \_\_\_\_\_ is/are objected to. 8) Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement. Application Papers 9) The specification is objected to by the Examiner. 10) The drawing(s) filed on is/are; a) accepted or b) objected to by the Examiner. Applicant may not request that any objection to the drawing(s) be held in abevance. See 37 CFR 1.85(a). Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d). 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152. Priority under 35 U.S.C. § 119 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some \* c) None of: Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### **DETAILED ACTION**

### Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

The factual inquiries set forth in *Graham* v. *John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

- Determining the scope and contents of the prior art.
- 2. Ascertaining the differences between the prior art and the claims at issue.
- Resolving the level of ordinary skill in the pertinent art.
- Considering objective evidence present in the application indicating obviousness or nonobviousness.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Application/Control Number: 10/785,082

Art Unit: 1797

Claims 49-51, 53-56 and 58-70 are rejected under 35 U.S.C. 103(a) as being unpatentable over Earley et al. (USPN 5.455,008).

Earley teaches a system for manipulating and thermal cycling a sample well tray, comprising: a thermal cycling device (fig. 1, ref. 14) having sample blocks (14, fig. 5) for receiving the sample well tray (fig. 7, ref. 34), the sample block comprising a plurality of openings (fig. 5, ref. 31) for receiving sample wells of the sample well tray; and a sample well tray handling apparatus comprising; a sample well tray holder (fig. 2, ref. 7 configured to receive the sample well tray therein; a rotational actuator configured to rotate the sample well tray holder about a rotational axis (col. 3, lines 1-col. 4, line 2): and an extension arm extending from the rotational actuator and being connected to the sample well tray holder (col. 3, lines 1-col. 4, line 2), wherein the sample well tray handling apparatus is configured to move the sample well tray from a first location distal to the sample block to a second location proximal to the sample block and then later remove the sample well tray from the second location to the first location (col. 3, lines 1col. 4, line 2). Earley states the microtiter plate is filled with a sample then placed on a heater 14 for five minutes at 65 degrees C to complete the denaturation step. The microtiter plate is then removed and placed on the workstation 16 where it is held at room temperature for five minutes to complete the annealing step. Next reagents are added to the microtiter plate and the plate is placed back onto the heating block 14 to begin the extension step at 65 degrees C.

Earley does not specifically teach employing only one heater block.

Earley employs multiple heating blocks 14 to in order to process large amounts of samples but states that each heating block performs the sample heating step as the next. Therefore it would have been obvious to one having an ordinary skill in the art at the time of the invention to modify Earley to reduce the number of heater blocks to one in order to process one microtiter plate at a time. This omission of multiple heaters with the consequent loss of its function is well known in the art (In re Kuhle, 188 U.S.P.Q. 7) in order to reduce costs and provide a simpler device.

# Double Patenting

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 49-51, 53-56 and 58-70 are rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claim 6,719,949 of U.S. Patent No. 1-22. Although the conflicting claims are not identical, they are not patentably distinct from each other because the instant claims are a broader version of the 6,719,449 Patent. The prior art claims include all the limitations found in the presently pending claims. Thus, the presently pending claims cover all the subject matter of the prior art 6,719,449 Patent.

### Response to Arguments

Applicant's arguments with respect to claims 49-51, 53-56 and 58-70 have been considered but are moot in view of the new ground(s) of rejection.

#### Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to SAM P. SIEFKE whose telephone number is (571)272-1262. The examiner can normally be reached on M-F 9:00am-5:30pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on 571-272-1700. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Samuel P Siefke/ Primary Examiner, Art Unit 1797